

MOST IMMEDIATE

I – 34 / 4 / 05 – FIN

Government of India

Office of the Director General of Civil Aviation

Opposite Safdarjung Airport, New Delhi – 110 003

.....

Dated the 07.03.2006.

TENDER NOTICE

Sealed quotations in the prescribed format are invited from the Original Equipment Manufacturers / Suppliers for supply of two Treadmill Exercise ECG Machine(TMT) and two Ambulatory BP Machine of the enclosed specifications one each to Air Force Central Medical Establishment, Subroto Park, New Delhi and IAM, Bangalore. The tender document may also be downloaded from website

<http://www.dgca.nic.in>

2. The quotations should be submitted separately for each equipment in two envelopes superscribed as Technical Bid for Ambulatory B.P. Machine and Technical Bid for Tread Mill Exercise Machine, Price Bids for Ambulatory B.P. Machine and Price Bid for Tread Mill Exercise Machine by 1400 hours on March 20, 2006 addressed to Deputy Director (Administration), Office of DGCA, Opposite Safdarjung Airport, New Delhi – 110 003. The tenders will be opened on **the same day at 3:00 P.M.** in the presence of **authorised representatives of the firms who wish to participate.** The competent authority reserves the right to accept or reject any or all the tenders without assigning any reason thereof.

3. The bids should be accompanied by an Earnest Money Deposit of Rs.40,000 (Rupees forty thousand only) covering both proposals in the form of Demand Draft/Banker's Cheque in favour of Pay & Accounts Officer, Central Pay & Accounts Office, Civil Aviation Department, New Delhi. Bids not accompanied by the required EMD shall be rejected outright.

(R.S. Kataria)

Deputy Director of Administration
for Director General of Civil Aviation

Encl: Format for Financial Bid and Technical Bid for both the equipments.

TECHNICAL BID for Treadmill Exercise ECG Machine (TMT)

Sl. No.	Specification	Whether the Requirements have been met. Yes / No.
1.	System should acquire/analys 12/15 leads	
2.	System should have wireless transmission of patient ECG to analyser	
3.	Should provide standard full interpretation of supine ECG with reasoning	
4.	Display of real time 12 lead diagnostic quality ECG waveform, average complexes beat of all 12 leads with colour comparison along with digital value level and slope	
5.	Display of four times enlarged complex with colour coded comparison with reference and facility for dynamic lead selection for maximum ST changes	
6.	System should display running trend of ST, Heart rate and METs on the screen to evaluate patient condition during exercise phase	
7.	System should have an ability to readjust J-ST interval measurement points and generate re-analysed new reports multiple patients data stored in system	
8.	Should have full disclosure playback review and storage for minimum of 50 tests	
9.	Reports should be printed in plain paper on Laser printer and optional thermal printer	
10.	Compatible programmable Non-Invasive blood pressure measurement module with bi-directional transfer of data between module and analyser	
11.	Equipment should comply FDA/CE rules and regulations	
12.	Must be ISO certified	
13.	UPS support for the entire system	
14.	Warranty period	
15.	Whether the specified EMD is enclosed	

Signature of the Contractor

TECHNICAL BID for Ambulatory B.P. Machine

Sl. No .	Specification	Whether the Requirements have been met. Yes / No.
1.	Should work on oscillometric technology for quick hook up, making it easy to use.	
2.	BP data should be stored on flash memory card means no loss of data	
3.	Should work with two 1.5 Volts AA Alkaline or rechargeable batteries for upto 52 hours or 250+BP readings.	
4.	Should weigh less than 300 gms including batteries.	
5.	Should generate concise clinical reports including full disclosure of BP readings, heart rate, mean arterial pressure, BP load and trend of BP value verses time.	
6.	Should work on Windows based PC software for Monitor Programming with comprehensive reports and editing capabilities.	
7.	ISO Certificate	
8.	Whether specified EMD is enclosed	

Signature of the Contractor

FINANCIAL BID for Treadmill Exercise ECG Machine (TMT)

- (a) Cost of the one unit of Treadmill Exercise ECG Machine :
 - (b) Taxes and other levies
 - (c) To indicate whether (a) includes cost of delivery
-
-

FINANCIAL BID for Ambulatory B.P. Machine

- (a) Cost of one unit of the Ambulatory B.P. Machine:
- (b) Taxes & Other levies :
- (c) To indicate whether (a) includes cost of delivery